



Near North Laboratories Inc.

UNIT 11 - 191 BOOTH RD, NORTH BAY ON P1A4K3

TEL (705) 497 0550 FAX (705) 497 0549

nearnorthlabs.ca

Sampling Instructions for CHEMICAL Testing (Drinking Water)

(including Reg 170/03; version 5, Nov 29, 2018):

SAMPLING INSTRUCTIONS

- 1) **Read instructions thoroughly before proceeding.**
- 2) Fill out each bottle label and the submission form completely.
- 3) Choose a sampling location that accurately represents the quality of the water being sampled. When sampling distribution/plumbing system taps, it is advisable not to sample from an unprotected outside tap.
- 4) Remove any aerators, hose attachments, filters and strainers from taps.
- 5) Open the **COLD** water tap and flush the line long enough to ensure fresh water is in the line (generally 5 to 10 minutes).
- 6) Pre-rinse each bottle with the sample to be taken.(x3) prior to collecting sample.
EXCEPTION: DO NOT PRE-RINSE Septum Vials, 1L Clear Glass(Oil & Grease), Or any bottles containing a Preservative as indicated on bottle.
- 7) Fill the bottles with sample to the shoulder leaving approximately one inch of headspace.
EXCEPTION: There should be NO AIR space in septum vials. Do not touch the inside of the cap or the lip of the bottle -- the inside of the cap should not come in contact with anything other than the atmosphere and the sample.
- 8) Replace the cap tightly. (Any required Preservation is done at Laboratory at sample reception)
- 9) Keep the sample cold (between 2 and 10 °C -- do not freeze) Transport samples to the lab in a rigid walled sealed box or cooler. Transport to the lab ASAP and within 24 hours (ideally the same day).

If you have any questions, please call the laboratory before proceeding with the sampling.

Reference: 4464e MOE Practices for the Collection and Handling of Drinking-Water Samples April 1,2009